

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100920\
 Data File : FE032949.D
 Signal(s) : FID1B.CH
 Acq On : 09 Oct 2020 23:03
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Oct 09 23:51:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
 Quant Title :
 QLast Update : Tue Sep 15 03:21:18 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.674	15289594	48.801 ug/ml
Target Compounds			
2) N-DECANE	4.414	16761068	49.833 ug/ml
3) N-DODECANE	6.430	17182590	50.785 ug/ml
4) N-TETRADECANE	8.225	17349449	51.491 ug/ml
5) N-HEXADECANE	9.827	17754878	51.280 ug/ml
6) N-OCTADECANE	11.269	17853306	50.881 ug/ml
7) N-EICOSANE	12.579	17625346	50.404 ug/ml
8) N-DOCOSANE	13.777	17215487	49.457 ug/ml
10) N-TETRACOSANE	14.880	17176771	48.869 ug/ml
11) N-HEXACOSANE	15.901	16369371	46.944 ug/ml
12) N-OCTACOSANE	16.851	16061196	46.297 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE100920\
Data File : FE032949.D
Signal(s) : FID1B.CH
Acq On : 09 Oct 2020 23:03
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Oct 09 23:51:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE091420.M
Quant Title :
QLast Update : Tue Sep 15 03:21:18 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

