

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE101621\
 Data File : FE034453.D
 Signal(s) : FID1B.CH
 Acq On : 16 Oct 2021 4:09
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 16 04:41:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100821.M
 Quant Title :
 QLast Update : Fri Oct 08 16:08:59 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.624	16387314	48.057 ug/ml
Target Compounds			
2) N-DECANE	4.394	19129942	49.796 ug/ml
3) N-DODECANE	6.399	19322383	50.610 ug/ml
4) N-TETRADECANE	8.190	19471409	49.488 ug/ml
5) N-HEXADECANE	9.788	19240834	48.715 ug/ml
6) N-OCTADECANE	11.227	19430066	48.411 ug/ml
7) N-EICOSANE	12.533	19306067	47.985 ug/ml
8) N-DOCOSANE	13.729	18922004	48.147 ug/ml
10) N-TETRACOSANE	14.830	18720416	48.075 ug/ml
11) N-HEXACOSANE	15.848	18307140	47.988 ug/ml
12) N-OCTACOSANE	16.794	17885701	47.380 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE101621\
Data File : FE034453.D
Signal(s) : FID1B.CH
Acq On : 16 Oct 2021 4:09
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 16 04:41:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE100821.M
Quant Title :
QLast Update : Fri Oct 08 16:08:59 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

