

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE101718\
 Data File : FE028917.D
 Signal(s) : FID1B.CH
 Acq On : 17 Oct 2018 16:52
 Operator : DD\SJ
 Sample : J5524-05MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 18 08:16:24 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
 Quant Title :
 QLast Update : Fri Sep 21 07:24:23 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.926	3424968	16.583 ug/ml
Target Compounds			
1) N-OCTANE	2.529	1490953	15.021 ug/ml
3) N-DECANE	4.621	2757386	17.230 ug/ml
5) N-DODECANE	6.647	3388886	27.932 ug/ml
6) N-TETRADECANE	8.451	3921450	29.565 ug/ml
7) N-HEXADECANE	10.058	4152530	27.743 ug/ml
8) N-OCTADECANE	11.506	4098290	24.950 ug/ml
9) N-EICOSANE	12.821	3990574	22.753 ug/ml
10) N-TETRACOSANE	15.131	3819824	21.901 ug/ml
11) N-OCTACOSANE	17.109	4194912	27.317 ug/ml
12) N-DOTRIACONTANE	18.835	4178886	32.167 ug/ml
14) N-HEXATRIACONTANE	20.361	3036190	30.047 ug/ml
15) N-TETRACONTANE	22.153	2561836	35.431 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE101718\
Data File : FE028917.D
Signal(s) : FID1B.CH
Acq On : 17 Oct 2018 16:52
Operator : DD\SJ
Sample : J5524-05MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 18 08:16:24 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE092118.M
Quant Title :
QLast Update : Fri Sep 21 07:24:23 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

