

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE101920\
 Data File : FE033017.D
 Signal(s) : FID1B.CH
 Acq On : 19 Oct 2020 10:20
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 20 00:49:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE101420.M
 Quant Title :
 QLast Update : Wed Oct 14 02:47:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.669	15297836	49.397 ug/ml
Target Compounds			
2) N-DECANE	4.411	17836152	51.733 ug/ml
3) N-DODECANE	6.425	17965154	52.117 ug/ml
4) N-TETRADECANE	8.219	17934467	51.900 ug/ml
5) N-HEXADECANE	9.821	18269376	51.501 ug/ml
6) N-OCTADECANE	11.263	18310903	50.779 ug/ml
7) N-EICOSANE	12.573	18035598	50.268 ug/ml
8) N-DOCOSANE	13.771	17710402	49.789 ug/ml
10) N-TETRACOSANE	14.876	17562534	49.283 ug/ml
11) N-HEXACOSANE	15.897	17050384	48.665 ug/ml
12) N-OCTACOSANE	16.847	16544612	47.739 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE101920\
Data File : FE033017.D
Signal(s) : FID1B.CH
Acq On : 19 Oct 2020 10:20
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 20 00:49:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE101420.M
Quant Title :
QLast Update : Wed Oct 14 02:47:05 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

