

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE102318\
 Data File : FE028947.D
 Signal(s) : FID1B.CH
 Acq On : 23 Oct 2018 20:01
 Operator : DD\SJ
 Sample : 1760 TRPH STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: sample.e
 Quant Time: Oct 24 08:16:28 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE102318.M
 Quant Title :
 QLast Update : Wed Oct 24 08:15:18 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.916	3258708	18.154 ug/ml
Target Compounds			
1) N-OCTANE	2.534	17566997	129.919 ug/ml
2) N-NONANE	3.555	17985207	131.003 ug/ml
3) N-DECANE	4.633	27491400	132.192 ug/ml
4) N-UNDECANE	5.674	29377001	199.981 ug/mlm
5) N-DODECANE	6.659	29788835	203.471 ug/ml
6) N-TETRADECANE	8.465	30599473	207.703 ug/ml
7) N-HEXADECANE	10.072	27112217	177.134 ug/ml
8) N-OCTADECANE	11.510	14169070	90.986 ug/ml
9) N-EICOSANE	12.815	5732317	36.649 ug/ml
10) N-TETRACOSANE	15.125	5534722	36.060 ug/ml
11) N-OCTACOSANE	17.098	2595937	17.649 ug/ml
12) N-DOTRIACONTANE	18.823	2233716	17.405 ug/ml
14) N-HEXATRIACONTANE	20.350	1819932	17.589 ug/ml
15) N-TETRACONTANE	22.130	1567421	17.469 ug/ml
16) N-TETRATETRAOCTANE	25.182	1351158	18.024 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE102318\
Data File : FE028947.D
Signal(s) : FID1B.CH
Acq On : 23 Oct 2018 20:01
Operator : DD\SJ
Sample : 1760 TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: sample.e
Quant Time: Oct 24 08:16:28 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE102318.M
Quant Title :
QLast Update : Wed Oct 24 08:15:18 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

