

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE102320\
 Data File : FE033044.D
 Signal(s) : FID1B.CH
 Acq On : 23 Oct 2020 9:48
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Ankita
 10/27/2020 10:04:35 AM

Integration File: autoint1.e
 Quant Time: Oct 23 23:59:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE101420.M
 Quant Title :
 QLast Update : Wed Oct 14 02:47:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.657	16551398	53.444 ug/ml
Target Compounds			
1) N-OCTANE	2.347	17726647	51.593 ug/ml
2) N-DECANE	4.404	19195577	55.676 ug/ml
3) N-DODECANE	6.418	19403985	56.292 ug/ml
4) N-TETRADECANE	8.211	19372343	56.061 ug/ml
5) N-HEXADECANE	9.811	19721737	55.595 ug/ml
6) N-OCTADECANE	11.252	19754823	54.783 ug/ml
7) N-EICOSANE	12.562	19448319	54.206 ug/ml
8) N-DOCOSANE	13.760	19039044	53.525 ug/ml
10) N-TETRACOSANE	14.864	18886059	52.997 ug/ml
11) N-HEXACOSANE	15.885	18296819	52.223 ug/ml
12) N-OCTACOSANE	16.835	17787670	51.326 ug/ml
13) N-TRIACONTANE	17.724	17441541	50.989 ug/ml
14) N-DOTRIACONTANE	18.555	16574428	51.206 ug/ml
15) N-TETRATRIACONTANE	19.338	15166441	51.931 ug/ml
16) N-HEXATRIACONTANE	20.076	13397250	52.852 ug/mlm
17) N-OCTATRIACONTANE	20.817	12201802	54.031 ug/ml
18) N-TETRACONTANE	21.735	11967580	54.788 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE102320\
Data File : FE033044.D
Signal(s) : FID1B.CH
Acq On : 23 Oct 2020 9:48
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
10/27/2020 10:04:35 AM

Integration File: autoint1.e
Quant Time: Oct 23 23:59:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE101420.M
Quant Title :
QLast Update : Wed Oct 14 02:47:05 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

