

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_E\Data\FE103018\
 Data File : FE028980.D
 Signal(s) : FID1B.CH
 Acq On : 30 Oct 2018 15:23
 Operator : DD\SJ
 Sample : J5697-01MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 31 01:05:53 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE102318.M
 Quant Title :
 QLast Update : Wed Oct 24 08:22:18 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.907	3509718	20.890 ug/ml
Target Compounds			
1) N-OCTANE	2.529	1796247	15.508 ug/ml
3) N-DECANE	4.607	2781198	15.489 ug/ml
5) N-DODECANE	6.630	3183467	24.667 ug/ml
6) N-TETRADECANE	8.503	1259276	9.503 ug/ml
7) N-HEXADECANE	10.039	3553326	25.179 ug/ml
8) N-OCTADECANE	11.487	3649008	24.834 ug/ml
9) N-EICOSANE	12.802	3430296	23.153 ug/ml
10) N-TETRACOSANE	15.112	4486209	31.392 ug/ml
11) N-OCTACOSANE	17.089	3486131	25.619 ug/ml
12) N-DOTRIACONTANE	18.814	3213209	26.961 ug/ml
14) N-HEXATRIACONTANE	20.341	2692008	27.895 ug/ml
15) N-TETRACONTANE	22.118	2543958	29.473 ug/ml
16) N-TETRTETRACOCTANE	25.144	176207	2.253 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_E\Data\FE103018\
Data File : FE028980.D
Signal(s) : FID1B.CH
Acq On : 30 Oct 2018 15:23
Operator : DD\SJ
Sample : J5697-01MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 31 01:05:53 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE102318.M
Quant Title :
QLast Update : Wed Oct 24 08:22:18 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

