

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103120\
 Data File : FE033091.D
 Signal(s) : FID1B.CH
 Acq On : 30 Oct 2020 14:08
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 31 00:51:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE101420.M
 Quant Title :
 QLast Update : Wed Oct 14 02:47:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.641	15711918	50.734 ug/ml
Target Compounds			
2) N-DECANE	4.395	18566470	53.851 ug/ml
3) N-DODECANE	6.406	18721381	54.311 ug/ml
4) N-TETRADECANE	8.198	18608122	53.850 ug/ml
5) N-HEXADECANE	9.798	18902146	53.285 ug/ml
6) N-OCTADECANE	11.239	18883339	52.367 ug/ml
7) N-EICOSANE	12.547	18562936	51.738 ug/ml
8) N-DOCOSANE	13.743	18121313	50.945 ug/ml
10) N-TETRACOSANE	14.848	17952718	50.378 ug/ml
11) N-HEXACOSANE	15.867	17341436	49.496 ug/ml
12) N-OCTACOSANE	16.816	16940412	48.881 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE103120\
Data File : FE033091.D
Signal(s) : FID1B.CH
Acq On : 30 Oct 2020 14:08
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 31 00:51:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE101420.M
Quant Title :
QLast Update : Wed Oct 14 02:47:05 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

