

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110320\
 Data File : FE033102.D
 Signal(s) : FID1B.CH
 Acq On : 03 Nov 2020 8:04
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Nov 04 00:14:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE110220.M
 Quant Title :
 QLast Update : Tue Nov 03 00:53:10 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.637	16669905	52.786 ug/ml
Target Compounds			
2) N-DECANE	4.393	18370369	54.232 ug/ml
3) N-DODECANE	6.402	18581927	55.150 ug/ml
4) N-TETRADECANE	8.193	18749150	54.972 ug/ml
5) N-HEXADECANE	9.792	18963660	54.336 ug/ml
6) N-OCTADECANE	11.234	19186771	53.602 ug/ml
7) N-EICOSANE	12.543	19099575	53.025 ug/ml
8) N-DOCOSANE	13.740	18873129	52.858 ug/ml
10) N-TETRACOSANE	14.843	18911092	52.765 ug/ml
11) N-HEXACOSANE	15.862	18608785	52.860 ug/ml
12) N-OCTACOSANE	16.813	18321773	52.648 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110320\
Data File : FE033102.D
Signal(s) : FID1B.CH
Acq On : 03 Nov 2020 8:04
Operator : DD\AJ
Sample : 50 TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Nov 04 00:14:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE110220.M
Quant Title :
QLast Update : Tue Nov 03 00:53:10 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

