

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE110524AL\  
Data File : FE051050.D  
Signal(s) : FID1B.ch  
Acq On : 06 Nov 2024 03:29  
Operator : YP\AJ  
Sample : P4701-06  
Misc :  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :

Integration File: autoint1.e  
Quant Time: Nov 06 04:31:16 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 100824.M  
Quant Title : GC Extractables  
QLast Update : Mon Oct 07 16:32:18 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.853 | 7027570    | 44.575 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 89.15%       |
| 12) S 1-chlorooctadecane (S... | 13.285 | 5838061    | 49.352 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 98.70%       |

Target Compounds

-----

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE110524AL\  
Data File : FE051050.D  
Signal(s) : FID1B.ch  
Acq On : 06 Nov 2024 03:29  
Operator : YP\AJ  
Sample : P4701-06  
Misc :  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :

Integration File: autoint1.e  
Quant Time: Nov 06 04:31:16 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 100824.M  
Quant Title : GC Extractables  
QLast Update : Mon Oct 07 16:32:18 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.853 min  
Delta R.T.: -0.010 min  
Response: 7027570  
Conc: 44.58 ug/ml

Instrument :  
FID\_E  
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 13.285 min  
Delta R.T.: -0.013 min  
Response: 5838061  
Conc: 49.35 ug/ml