

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110618\
 Data File : FE029018.D
 Signal(s) : FID1B.CH
 Acq On : 06 Nov 2018 10:32
 Operator : DD\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 07 08:35:14 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE102518.M
 Quant Title :
 QLast Update : Thu Oct 25 04:58:05 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.928	8065562	46.873 ug/ml
Target Compounds			
2) N-DECANE	4.614	9693349	58.777 ug/ml
3) N-DODECANE	6.641	10261528	59.405 ug/ml
4) N-TETRADECANE	8.445	9804544	53.083 ug/ml
5) N-HEXADECANE	10.055	9749408	50.288 ug/ml
6) N-OCTADECANE	11.505	9609695	47.363 ug/ml
7) N-EICOSANE	12.821	9423039	46.833 ug/ml
8) N-DOCOSANE	14.026	9572384	47.715 ug/ml
10) N-TETRACOSANE	15.136	9108953	46.915 ug/ml
11) N-HEXACOSANE	16.161	9147883	48.573 ug/ml
12) N-OCTACOSANE	17.117	9099496	49.904 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110618\
Data File : FE029018.D
Signal(s) : FID1B.CH
Acq On : 06 Nov 2018 10:32
Operator : DD\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 07 08:35:14 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE102518.M
Quant Title :
QLast Update : Thu Oct 25 04:58:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

