

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110618\
 Data File : FE029021.D
 Signal(s) : FID1B.CH
 Acq On : 06 Nov 2018 14:04
 Operator : DD\SJ
 Sample : PB114546BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 07 08:35:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE102518.M
 Quant Title :
 QLast Update : Thu Oct 25 04:58:05 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.918	3227325	18.756 ug/ml
Target Compounds			
2) N-DECANE	4.607	3897606	23.634 ug/mlm
3) N-DODECANE	6.633	4174189	24.165 ug/ml
4) N-TETRADECANE	8.437	4242992	22.972 ug/ml
5) N-HEXADECANE	10.046	4255440	21.950 ug/ml
6) N-OCTADECANE	11.495	4172864	20.567 ug/ml
7) N-EICOSANE	12.812	4166445	20.707 ug/ml
8) N-DOCOSANE	14.017	4067071	20.273 ug/ml
10) N-TETRACOSANE	15.125	4020724	20.708 ug/ml
11) N-HEXACOSANE	16.152	3949796	20.973 ug/mlm
12) N-OCTACOSANE	17.106	3940125	21.609 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110618\
Data File : FE029021.D
Signal(s) : FID1B.CH
Acq On : 06 Nov 2018 14:04
Operator : DD\SJ
Sample : PB114546BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 07 08:35:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE102518.M
Quant Title :
QLast Update : Thu Oct 25 04:58:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

