

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110620\
 Data File : FE033120.D
 Signal(s) : FID1B.CH
 Acq On : 07 Nov 2020 1:05
 Operator : DD\AJ
 Sample : PB132862BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 09 16:45:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE110620.M
 Quant Title :
 QLast Update : Mon Nov 09 16:39:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.622	6163792	20.502 ug/ml
Target Compounds			
1) N-OCTANE	2.350	6281671	21.662 ug/ml
3) N-DECANE	4.384	6704456	21.708 ug/ml
5) N-DODECANE	6.391	6903369	22.093 ug/ml
6) N-TETRADECANE	8.182	7099303	22.664 ug/ml
7) N-HEXADECANE	9.780	7173117	23.089 ug/ml
8) N-OCTADECANE	11.219	7256524	23.586 ug/ml
9) N-EICOSANE	12.527	7375633	24.602 ug/ml
10) N-TETRACOSANE	14.825	7219337	24.268 ug/ml
11) N-OCTACOSANE	16.793	6946718	24.417 ug/ml
12) N-DOTRIACONTANE	18.514	6547595	25.096 ug/ml
14) N-HEXATRIACONTANE	20.036	5579614	24.851 ug/ml
15) N-TETRACONTANE	21.673	5038463	24.421 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110620\
Data File : FE033120.D
Signal(s) : FID1B.CH
Acq On : 07 Nov 2020 1:05
Operator : DD\AJ
Sample : PB132862BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 09 16:45:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE110620.M
Quant Title :
QLast Update : Mon Nov 09 16:39:48 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

