

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110719\
 Data File : FE030783.D
 Signal(s) : FID1B.CH
 Acq On : 07 Nov 2019 18:18
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Ankita
 11/8/2019 8:38:49 AM

Integration File: autoint1.e
 Quant Time: Nov 08 00:33:21 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
 Quant Title :
 QLast Update : Thu Oct 31 08:47:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.689	15793727	50.818 ug/ml
Target Compounds			
2) N-DECANE	4.413	19981957	52.515 ug/ml
3) N-DODECANE	6.427	19960364	52.459 ug/ml
4) N-TETRADECANE	8.226	19626069	52.410 ug/ml
5) N-HEXADECANE	9.830	19465910	52.192 ug/ml
6) N-OCTADECANE	11.274	19003271	51.973 ug/ml
7) N-EICOSANE	12.588	18807437	51.461 ug/ml
8) N-DOCOSANE	13.789	18398248	51.177 ug/ml
10) N-TETRACOSANE	14.896	17976571	50.539 ug/ml
11) N-HEXACOSANE	15.917	17658205	50.633 ug/ml
12) N-OCTACOSANE	16.868	17210643	50.481 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE110719\
Data File : FE030783.D
Signal(s) : FID1B.CH
Acq On : 07 Nov 2019 18:18
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
11/8/2019 8:38:49 AM

Integration File: autoint1.e
Quant Time: Nov 08 00:33:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
Quant Title :
QLast Update : Thu Oct 31 08:47:02 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

