

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111219\
 Data File : FE030836.D
 Signal(s) : FID1B.CH
 Acq On : 12 Nov 2019 19:33
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 13 00:43:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
 Quant Title :
 QLast Update : Thu Oct 31 08:47:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.672	14718324	47.358 ug/ml
Target Compounds			
2) N-DECANE	4.401	18351983	48.231 ug/ml
3) N-DODECANE	6.414	18355694	48.241 ug/ml
4) N-TETRADECANE	8.211	18072977	48.263 ug/ml
5) N-HEXADECANE	9.815	17983220	48.217 ug/ml
6) N-OCTADECANE	11.258	17657047	48.291 ug/ml
7) N-EICOSANE	12.571	17476116	47.818 ug/ml
8) N-DOCOSANE	13.772	17089513	47.537 ug/ml
10) N-TETRACOSANE	14.878	16717537	46.999 ug/ml
11) N-HEXACOSANE	15.900	16384188	46.980 ug/ml
12) N-OCTACOSANE	16.852	15965467	46.828 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111219\
Data File : FE030836.D
Signal(s) : FID1B.CH
Acq On : 12 Nov 2019 19:33
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 13 00:43:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
Quant Title :
QLast Update : Thu Oct 31 08:47:02 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

