

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111419\
 Data File : FE030859.D
 Signal(s) : FID1B.CH
 Acq On : 14 Nov 2019 16:24
 Operator : DD\AJ
 Sample : PB124777BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 15 00:11:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
 Quant Title :
 QLast Update : Thu Oct 31 08:47:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.660	5289639	17.020 ug/ml
Target Compounds			
2) N-DECANE	4.394	6512650	17.116 ug/ml
3) N-DODECANE	6.405	6550354	17.215 ug/ml
4) N-TETRADECANE	8.201	6472994	17.286 ug/ml
5) N-HEXADECANE	9.804	6435081	17.254 ug/ml
6) N-OCTADECANE	11.248	6329856	17.312 ug/ml
7) N-EICOSANE	12.560	6244037	17.085 ug/ml
8) N-DOCOSANE	13.759	6127198	17.044 ug/ml
10) N-TETRACOSANE	14.865	6017118	16.916 ug/ml
11) N-HEXACOSANE	15.887	5908669	16.943 ug/ml
12) N-OCTACOSANE	16.838	5875582	17.234 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111419\
Data File : FE030859.D
Signal(s) : FID1B.CH
Acq On : 14 Nov 2019 16:24
Operator : DD\AJ
Sample : PB124777BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 15 00:11:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
Quant Title :
QLast Update : Thu Oct 31 08:47:02 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

