

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111419\
 Data File : FE030861.D
 Signal(s) : FID1B.CH
 Acq On : 14 Nov 2019 17:31
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 15 00:11:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
 Quant Title :
 QLast Update : Thu Oct 31 08:47:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.666	15362300	49.430 ug/ml
Target Compounds			
2) N-DECANE	4.397	18835619	49.502 ug/ml
3) N-DODECANE	6.411	18933844	49.761 ug/ml
4) N-TETRADECANE	8.207	18722955	49.998 ug/ml
5) N-HEXADECANE	9.811	18666841	50.050 ug/ml
6) N-OCTADECANE	11.254	18307617	50.070 ug/ml
7) N-EICOSANE	12.566	18220356	49.855 ug/ml
8) N-DOCOSANE	13.766	17776597	49.448 ug/ml
10) N-TETRACOSANE	14.872	17421954	48.980 ug/ml
11) N-HEXACOSANE	15.894	17067392	48.939 ug/ml
12) N-OCTACOSANE	16.845	16657535	48.858 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111419\
Data File : FE030861.D
Signal(s) : FID1B.CH
Acq On : 14 Nov 2019 17:31
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 15 00:11:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
Quant Title :
QLast Update : Thu Oct 31 08:47:02 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

