

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111522AL\
 Data File : FE038783.D
 Signal(s) : FID1B.ch
 Acq On : 15 Nov 2022 19:47
 Operator : YP\AJ
 Sample : N5617-01 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 72-12013

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/16/2022
 Supervised By :Ankita Jodhani 11/16/2022

Integration File: autoint1.e
 Quant Time: Nov 16 04:03:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 102522.M
 Quant Title : GC Extractables
 QLast Update : Wed Oct 26 04:49:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.960	372088	2.162 ug/mlm
Spiked Amount 50.000		Recovery =	4.32%
12) S 1-chlorooctadecane (S...	13.396	275941	2.016 ug/ml
Spiked Amount 50.000		Recovery =	4.03%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111522AL\
Data File : FE038783.D
Signal(s) : FID1B.ch
Acq On : 15 Nov 2022 19:47
Operator : YP\AJ
Sample : N5617-01 10X
Misc :
ALS Vial : 17 Sample Multiplier: 1


Instrument :
FID_E
ClientSampleId :
72-12013


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/16/2022
Supervised By :Ankita Jodhani 11/16/2022

Integration File: autoint1.e
Quant Time: Nov 16 04:03:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_D\methods\Aliphatic EPH 102522.M
Quant Title : GC Extractables
QLast Update : Wed Oct 26 04:49:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.960 min
Delta R.T.: -0.036 min
Response: 372088
Conc: 2.16 ug/ml

Instrument :
FID_E
ClientSampleId :
72-12013

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/16/2022
Supervised By :Ankita Jodhani 11/16/2022



#12 1-chlorooctadecane (SURR)

R.T.: 13.396 min
Delta R.T.: -0.030 min
Response: 275941
Conc: 2.02 ug/ml