

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111618\
 Data File : FE029256.D
 Signal(s) : FID1B.CH
 Acq On : 16 Nov 2018 13:22
 Operator : AJ\SJ
 Sample : 440 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 16 14:45:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE111618.M
 Quant Title :
 QLast Update : Fri Nov 16 14:42:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.852	921464	5.149 ug/ml
Target Compounds			
1) N-OCTANE	2.486	5720861	43.347 ug/ml
2) N-NONANE	3.496	5808614	42.060 ug/ml
3) N-DECANE	4.565	8880196	41.427 ug/mlm
4) N-UNDECANE	5.603	9494428	61.598 ug/mlm
5) N-DODECANE	6.587	9690296	61.933 ug/ml
6) N-TETRADECANE	8.391	10204133	62.751 ug/ml
7) N-HEXADECANE	9.997	9380027	54.170 ug/ml
8) N-OCTADECANE	11.438	4925908	27.575 ug/ml
9) N-EICOSANE	12.748	1905569	10.573 ug/ml
10) N-TETRACOSANE	15.056	1843145	10.576 ug/ml
11) N-OCTACOSANE	17.033	823083	5.211 ug/ml
12) N-DOTRIACONTANE	18.758	803341	5.585 ug/ml
14) N-HEXATRIACONTANE	20.285	779550	5.356 ug/ml
15) N-TETRACONTANE	22.026	757125	5.353 ug/ml
16) N-TETRATETRAOCTANE	24.986	677221	5.587 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE111618\
Data File : FE029256.D
Signal(s) : FID1B.CH
Acq On : 16 Nov 2018 13:22
Operator : AJ\SJ
Sample : 440 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 16 14:45:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE111618.M
Quant Title :
QLast Update : Fri Nov 16 14:42:03 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

