

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112118\
 Data File : FE029283.D
 Signal(s) : FID1B.CH
 Acq On : 21 Nov 2018 11:20
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 21 11:49:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE111618.M
 Quant Title :
 QLast Update : Fri Nov 16 14:56:57 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.854	8374487	45.375 ug/ml
Target Compounds			
1) N-OCTANE	2.499	52238181	377.543 ug/ml
2) N-NONANE	3.517	55329966	386.179 ug/ml
3) N-DECANE	4.598	87034683	393.838 ug/ml
4) N-UNDECANE	5.640	94363939	594.630 ug/ml
5) N-DODECANE	6.626	96113790	593.382 ug/ml
6) N-TETRADECANE	8.432	99515684	589.655 ug/ml
7) N-HEXADECANE	10.035	86909415	478.716 ug/ml
8) N-OCTADECANE	11.463	43713039	232.596 ug/ml
9) N-EICOSANE	12.760	17285346	92.537 ug/ml
10) N-TETRACOSANE	15.069	16273030	90.159 ug/ml
11) N-OCTACOSANE	17.034	7597960	46.812 ug/ml
12) N-DOTRIACONTANE	18.758	7343908	48.440 ug/ml
14) N-HEXATRIACONTANE	20.286	7441914	48.842 ug/ml
15) N-TETRACONTANE	22.030	7699229	51.101 ug/ml
16) N-TETRATETRAOCTANE	24.987	6874507	50.447 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112118\
Data File : FE029283.D
Signal(s) : FID1B.CH
Acq On : 21 Nov 2018 11:20
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 21 11:49:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE111618.M
Quant Title :
QLast Update : Fri Nov 16 14:56:57 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

