

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112219\
 Data File : FE030928.D
 Signal(s) : FID1B.CH
 Acq On : 22 Nov 2019 11:06
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 22 21:07:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
 Quant Title :
 QLast Update : Thu Oct 31 08:47:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.668	16365857	52.659 ug/ml
Target Compounds			
2) N-DECANE	4.394	19773048	51.966 ug/ml
3) N-DODECANE	6.408	19988754	52.533 ug/ml
4) N-TETRADECANE	8.206	19744650	52.727 ug/ml
5) N-HEXADECANE	9.810	19711016	52.850 ug/ml
6) N-OCTADECANE	11.254	19367627	52.969 ug/ml
7) N-EICOSANE	12.567	19269570	52.726 ug/ml
8) N-DOCOSANE	13.768	18794961	52.281 ug/ml
10) N-TETRACOSANE	14.875	18459926	51.898 ug/ml
11) N-HEXACOSANE	15.898	18106137	51.918 ug/ml
12) N-OCTACOSANE	16.848	17811588	52.243 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112219\
Data File : FE030928.D
Signal(s) : FID1B.CH
Acq On : 22 Nov 2019 11:06
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 22 21:07:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE103019.M
Quant Title :
QLast Update : Thu Oct 31 08:47:02 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

