

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112320\
 Data File : FE033161.D
 Signal(s) : FID1B.CH
 Acq On : 23 Nov 2020 12:28
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Nov 24 00:10:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE110220.M
 Quant Title :
 QLast Update : Tue Nov 03 00:53:10 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.625	17596075	55.719 ug/ml
Target Compounds			
2) N-DECANE	4.383	20742835	61.235 ug/ml
3) N-DODECANE	6.392	21014131	62.369 ug/ml
4) N-TETRADECANE	8.184	21176312	62.089 ug/ml
5) N-HEXADECANE	9.783	20993398	60.151 ug/ml
6) N-OCTADECANE	11.223	21048502	58.803 ug/ml
7) N-EICOSANE	12.531	21159943	58.745 ug/ml
8) N-DOCOSANE	13.728	20830203	58.339 ug/ml
10) N-TETRACOSANE	14.831	20223032	56.426 ug/ml
11) N-HEXACOSANE	15.850	19588978	55.645 ug/ml
12) N-OCTACOSANE	16.800	19081982	54.833 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112320\
Data File : FE033161.D
Signal(s) : FID1B.CH
Acq On : 23 Nov 2020 12:28
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Nov 24 00:10:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE110220.M
Quant Title :
QLast Update : Tue Nov 03 00:53:10 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

