

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF112625AR\  
 Data File : FF016821.D  
 Signal(s) : FID2B.ch  
 Acq On : 26 Nov 2025 22:11  
 Operator : YP\AJ  
 Sample : 20 PPM AROMATIC HC STD3  
 Misc :  
 ALS Vial : 23 Sample Multiplier: 1

Instrument :  
 FID\_E  
 ClientSampleId :

Integration File: autoint1.e  
 Quant Time: Dec 01 08:12:00 2025  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\Aromatic EPH 112625.M  
 Quant Title : GC Extractables  
 QLast Update : Mon Dec 01 08:08:04 2025  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18µm

| Compound                           | R.T.   | Response   | Conc Units   |
|------------------------------------|--------|------------|--------------|
| -----                              |        |            |              |
| System Monitoring Compounds        |        |            |              |
| 4) S 2-Bromonaphthalene (S...      | 8.003  | 2377140    | 20.000 ug/ml |
| Spiked Amount 50.000               |        | Recovery = | 40.00%       |
| 6) S 2-Fluorobiphenyl (SURR)       | 8.879  | 1493215    | 20.000 ug/ml |
| Spiked Amount 50.000 Range 0 - 131 |        | Recovery = | 40.00%       |
| 11) S ortho-Terphenyl (SURR)       | 11.944 | 2974910    | 20.000 ug/ml |
| Spiked Amount 50.000               |        | Recovery = | 40.00%       |
| Target Compounds                   |        |            |              |
| 1) T 1,2,3-Trimethylbenzen...      | 4.597  | 2766830    | 20.000 ug/ml |
| 2) T Naphthalene (C11.7)           | 6.274  | 2796587    | 20.000 ug/ml |
| 3) T 2-Methylnaphthalene (...)     | 7.373  | 2846214    | 20.000 ug/ml |
| 5) T Acenaphthylene (C15.06)       | 8.678  | 2794456    | 20.000 ug/ml |
| 7) T Acenaphthene (C15.5)          | 8.977  | 2916628    | 20.000 ug/ml |
| 8) T Flouorene (C16.55)            | 9.768  | 2717253    | 20.000 ug/ml |
| 9) T Phenanthrene (C19.36)         | 11.175 | 2674025    | 20.000 ug/ml |
| 10) T Anthracene (C19.43)          | 11.250 | 2576963    | 20.000 ug/ml |
| 12) T Fluoranthene (C21.85)        | 12.998 | 2678897    | 20.000 ug/ml |
| 13) T Pyrene (C20.8)               | 13.296 | 2692958    | 20.000 ug/ml |
| 14) T Benzo[a]anthracene (C...     | 15.181 | 2255236    | 20.000 ug/ml |
| 15) T Chrysene (C27.41)            | 15.222 | 2758326    | 20.000 ug/ml |
| 16) T benzo[b]fluoranthene ...     | 16.744 | 2438049    | 20.000 ug/ml |
| 17) T Bnezo[k]fluoranthene ...     | 16.775 | 2587597    | 20.000 ug/ml |
| 18) T Benzo[a]pyrene (C31.34)      | 17.126 | 2469971    | 20.000 ug/ml |
| 19) T Indeno[1,2,3-cd]pyren...     | 18.504 | 2085274    | 20.077 ug/ml |
| 20) T Dibenz[a,h]anthracene...     | 18.541 | 3170642    | 19.954 ug/ml |
| 21) T Benzo[g,h,i]perylene ...     | 18.757 | 2622564    | 20.117 ug/ml |
| -----                              |        |            |              |

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_F\Data\FF112625AR\  
Data File : FF016821.D  
Signal(s) : FID2B.ch  
Acq On : 26 Nov 2025 22:11  
Operator : YP\AJ  
Sample : 20 PPM AROMATIC HC STD3  
Misc :  
ALS Vial : 23 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :

Integration File: autoint1.e  
Quant Time: Dec 01 08:12:00 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_F\Method\Aromatic EPH 112625.M  
Quant Title : GC Extractables  
QLast Update : Mon Dec 01 08:08:04 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18µm






#1 1,2,3-Trimethylbenzene (C10.1)


R.T.: 4.597 min  
Delta R.T.: 0.000 min  
Response: 2766830  
Conc: 20.00 ug/ml

Instrument :  
FID\_E  
ClientSampleId :




#2 Naphthalene (C11.7)

R.T.: 6.274 min  
Delta R.T.: 0.000 min  
Response: 2796587  
Conc: 20.00 ug/ml




#3 2-Methylnaphthalene (C12.89)

R.T.: 7.373 min  
Delta R.T.: 0.000 min  
Response: 2846214  
Conc: 20.00 ug/ml



#4 2-Bromonaphthalene (SURR)


R.T.: 8.003 min  
Delta R.T.: 0.000 min  
Response: 2377140  
Conc: 20.00 ug/ml



#5 Acenaphthylene (C15.06)


R.T.: 8.678 min  
Delta R.T.: 0.000 min  
Response: 2794456  
Conc: 20.00 ug/ml

Instrument :  
FID\_E  
ClientSampleId :




#6 2-Fluorobiphenyl (SURR)

R.T.: 8.879 min  
Delta R.T.: 0.000 min  
Response: 1493215  
Conc: 20.00 ug/ml




#7 Acenaphthene (C15.5)

R.T.: 8.977 min  
Delta R.T.: 0.000 min  
Response: 2916628  
Conc: 20.00 ug/ml



#8 Flouorene (C16.55)


R.T.: 9.768 min  
Delta R.T.: 0.000 min  
Response: 2717253  
Conc: 20.00 ug/ml



#9 Phenanthrene (C19.36)


R.T.: 11.175 min  
Delta R.T.: 0.000 min  
Response: 2674025  
Conc: 20.00 ug/ml

Instrument :  
FID\_E  
ClientSampleId :




#10 Anthracene (C19.43)

R.T.: 11.250 min  
Delta R.T.: 0.000 min  
Response: 2576963  
Conc: 20.00 ug/ml




#11 ortho-Terphenyl (SURR)

R.T.: 11.944 min  
Delta R.T.: 0.000 min  
Response: 2974910  
Conc: 20.00 ug/ml



#12 Fluoranthene (C21.85)


R.T.: 12.998 min  
Delta R.T.: 0.000 min  
Response: 2678897  
Conc: 20.00 ug/ml



#13 Pyrene (C20.8)


R.T.: 13.296 min  
Delta R.T.: 0.000 min  
Response: 2692958  
Conc: 20.00 ug/ml

Instrument :  
FID\_E  
ClientSampleId :




#14 Benzo[a]anthracene (C26.37)

R.T.: 15.181 min  
Delta R.T.: 0.000 min  
Response: 2255236  
Conc: 20.00 ug/ml




#15 Chrysene (C27.41)

R.T.: 15.222 min  
Delta R.T.: 0.000 min  
Response: 2758326  
Conc: 20.00 ug/ml



#16 benzo[b]fluoranthene (C30.41)


R.T.: 16.744 min  
Delta R.T.: 0.000 min  
Response: 2438049  
Conc: 20.00 ug/ml



#17 Bnezo[k]fluoranthene (C30.14)


R.T.: 16.775 min  
Delta R.T.: 0.000 min  
Response: 2587597  
Conc: 20.00 ug/ml

Instrument :  
FID\_E  
ClientSampleId :




#18 Benzo[a]pyrene (C31.34)

R.T.: 17.126 min  
Delta R.T.: 0.000 min  
Response: 2469971  
Conc: 20.00 ug/ml




#19 Indeno[1,2,3-cd]pyrene (C35.01)

R.T.: 18.504 min  
Delta R.T.: 0.000 min  
Response: 2085274  
Conc: 20.08 ug/ml



#20 Dibenz[a,h]anthracene (C30.36)

R.T.: 18.541 min  
Delta R.T.: 0.000 min  
Response: 3170642  
Conc: 19.95 ug/ml



#21 Benzo[g,h,i]perylene (C34.01)

R.T.: 18.757 min  
Delta R.T.: 0.000 min  
Response: 2622564  
Conc: 20.12 ug/ml

Instrument :

FID\_E

ClientSampleId :