

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112719\
 Data File : FE030970.D
 Signal(s) : FID1B.CH
 Acq On : 27 Nov 2019 17:13
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 27 21:39:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE112619.M
 Quant Title :
 QLast Update : Tue Nov 26 14:51:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.663	15008304	50.031 ug/ml
Target Compounds				
2)	N-DECANE	4.390	18837089	51.204 ug/ml
3)	N-DODECANE	6.402	18768698	51.124 ug/ml
4)	N-TETRADECANE	8.199	18451660	50.442 ug/ml
5)	N-HEXADECANE	9.803	18368076	50.235 ug/ml
6)	N-OCTADECANE	11.247	17985919	50.193 ug/ml
7)	N-EICOSANE	12.561	17948072	50.227 ug/ml
8)	N-DOCOSANE	13.762	17583039	50.079 ug/ml
10)	N-TETRACOSANE	14.869	17291710	49.440 ug/ml
11)	N-HEXACOSANE	15.892	17040240	49.749 ug/ml
12)	N-OCTACOSANE	16.845	16766050	49.363 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE112719\
Data File : FE030970.D
Signal(s) : FID1B.CH
Acq On : 27 Nov 2019 17:13
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 27 21:39:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE112619.M
Quant Title :
QLast Update : Tue Nov 26 14:51:59 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

