

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE120523AL\
 Data File : FE045651.D
 Signal(s) : FID1B.ch
 Acq On : 05 Dec 2023 12:47
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 05 15:32:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 113023.M
 Quant Title : GC Extractables
 QLast Update : Thu Nov 30 12:25:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.859	4124306	19.723 ug/ml
Spiked Amount 50.000		Recovery =	39.45%
12) S 1-chlorooctadecane (S...	13.302	3050536	19.387 ug/ml
Spiked Amount 50.000		Recovery =	38.77%
Target Compounds			
1) T n-Nonane (C9)	3.247	3279844	20.683 ug/ml
2) T n-Decane (C10)	4.498	3349565	20.917 ug/ml
3) T A~Naphthalene (C11.7)	6.197	3657219	20.995 ug/ml
4) T n-Dodecane (C12)	6.664	3388345	20.977 ug/ml
5) T A~2-methylnaphthalene...	7.293	3566187	21.000 ug/ml
6) T n-Tetradecane (C14)	8.491	3365850	20.981 ug/ml
7) T n-Hexadecane (C16)	10.101	3436105	20.750 ug/ml
8) T n-Octadecane (C18)	11.546	3476814	20.279 ug/ml
10) T n-Eicosane (C20)	12.858	3380700	19.539 ug/ml
11) T n-Heneicosane (C21)	13.470	3327943	19.345 ug/ml
13) T n-Docosane (C22)	14.057	3282251	19.036 ug/ml
14) T n-Tetracosane (C24)	15.162	3198518	18.321 ug/ml
15) T n-Hexacosane (C26)	16.185	3097782	17.788 ug/ml
16) T n-Octacosane (C28)	17.137	2997787	17.345 ug/ml
17) T n-Tricontane (C30)	18.026	3003655	17.169 ug/ml
18) T n-Dotriacontane (C32)	18.862	2972677	16.294 ug/ml
19) T n-Tetratriacontane (C34)	19.647	2853108	17.195 ug/ml
20) T n-Hexatriacontane (C36)	20.388	2751322	17.333 ug/ml
21) T n-Octatriacontane (C38)	21.122	2788076	17.779 ug/ml
22) T n-Tetracontane (C40)	22.027	2910854	20.482 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE120523AL\
Data File : FE045651.D
Signal(s) : FID1B.ch
Acq On : 05 Dec 2023 12:47
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 05 15:32:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 113023.M
Quant Title : GC Extractables
QLast Update : Thu Nov 30 12:25:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

