

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Rahul Chavli 12/09/2025
Supervised By :Yogesh Patel 12/09/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE12082
Data File : FE057212.D
Signal(s) : FID1B.ch
Acq On : 08 Dec 2025 09:34
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 112725.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.293	3.252	3.380	BB	390884	3892833	88.98%	4.735%
2	4.546	4.512	4.630	BB	403289	4034634	92.22%	4.907%
3	6.264	6.234	6.414	BB	342269	4056621	92.72%	4.934%
4	6.719	6.679	6.795	BB	408133	4105801	93.84%	4.994%
5	7.364	7.330	7.525	BB	321736	3828352	87.50%	4.656%
6	8.553	8.517	8.644	BB	383320	4043273	92.42%	4.918%
7	10.165	10.129	10.250	BB	377006	4110866	93.96%	5.000%
8	11.613	11.575	11.690	BB	353297	4099500	93.70%	4.986%
9	11.936	11.900	12.014	BB	396942	4375107	100.00%	5.321%
10	12.927	12.884	12.992	BB	337589	3880984	88.71%	4.720%
11	13.375	13.327	13.449	BB	273717	3261651	74.55%	3.967%
12	13.540	13.504	13.602	BB	311365	3796721	86.78%	4.618%
13	14.127	14.090	14.194	BB	299123	3738894	85.46%	4.547%
14	15.234	15.195	15.302	BB	289644	3675767	84.02%	4.471%
15	16.257	16.189	16.322	BB	287704	3610139	82.52%	4.391%
16	17.210	17.125	17.270	BB	279797	3579998	81.83%	4.354%
17	18.099	18.022	18.152	BB	254261	3700423	84.58%	4.501%
18	18.934	18.870	18.975	BB	251833	3633224	83.04%	4.419%
19	19.720	19.397	19.775	BV	238053	3386916	77.41%	4.119%
20	20.462	20.412	20.501	VV	224598	3205601	73.27%	3.899%
21	21.209	21.166	21.252	M	168915	3064557	70.05%	3.727%
22	22.140	22.050	22.220	BV	126583	3139329	71.75%	3.818%
Sum of corrected areas:							82221192	

Aliphatic EPH 112725.M Tue Dec 09 00:58:32 2025