

rteres

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report
Manual IntegrationsAPPROVED
Reviewed By :Yogesh Patel 12/13/2022
Supervised By :Ankita Jodhani 12/13/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE12122
Data File : FE039263.D
Signal(s) : FID1B.ch
Acq On : 13 Dec 2022 00:57
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 111922.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.334	3.298	3.382	BB	309076	3017699	82.94%	4.383%
2	4.583	4.540	4.630	BB	302612	2997168	82.37%	4.353%
3	6.285	6.246	6.367	BB	359435	3461920	95.15%	5.028%
4	6.746	6.703	6.794	BB	312991	3045179	83.69%	4.422%
5	7.379	7.338	7.467	BB	331132	3274486	90.00%	4.756%
6	8.571	8.527	8.628	BB	296925	3111780	85.52%	4.519%
7	10.178	10.135	10.232	BB	294740	3250113	89.33%	4.720%
8	11.620	11.582	11.672	BB	288034	3342542	91.87%	4.854%
9	11.938	11.903	11.995	BB	340419	3638487	100.00%	5.284%
10	12.929	12.893	12.977	BB	279315	3294028	90.53%	4.784%
11	13.372	13.338	13.425	BB	232510	2800182	76.96%	4.067%
12	13.539	13.505	13.591	M	274759	3314229	91.09%	4.813%
13	14.125	14.056	14.178	BB	269816	3293768	90.53%	4.784%
14	15.229	15.180	15.277	BB	263270	3263390	89.69%	4.739%
15	16.248	16.202	16.300	BB	238837	3189468	87.66%	4.632%
16	17.198	17.138	17.253	BB	230160	3170436	87.14%	4.604%
17	18.085	18.051	18.137	BB	231000	3177734	87.34%	4.615%
18	18.918	18.858	18.966	BB	224452	3167189	87.05%	4.600%
19	19.702	19.651	19.747	BB	214880	3001311	82.49%	4.359%
20	20.442	20.401	20.478	BV	184140	2727004	74.95%	3.960%
21	21.179	21.117	21.268	BB	154143	2745673	75.46%	3.988%
22	22.092	22.016	22.153	BB	112224	2572946	70.71%	3.737%
Sum of corrected areas:							68856730	

Aliphatic EPH 111922.M Tue Dec 13 05:00:05 2022