

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE121318\
 Data File : FE029330.D
 Signal(s) : FID1B.CH
 Acq On : 13 Dec 2018 19:09
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 14 00:40:32 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE120718.M
 Quant Title :
 QLast Update : Fri Dec 07 23:47:48 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.848	9802722	51.156 ug/ml
Target Compounds			
1) N-OCTANE	2.530	67256135	386.095 ug/ml
2) N-NONANE	3.534	71856519	412.353 ug/ml
3) N-DECANE	4.609	109828206	422.438 ug/ml
4) N-UNDECANE	5.646	115798217	637.782 ug/ml
5) N-DODECANE	6.631	114421384	635.126 ug/ml
6) N-TETRADECANE	8.433	112914059	631.675 ug/ml
7) N-HEXADECANE	10.035	96218298	520.747 ug/ml
8) N-OCTADECANE	11.461	48360651	261.066 ug/ml
9) N-EICOSANE	12.756	19358553	106.894 ug/ml
10) N-TETRACOSANE	15.064	19123414	102.840 ug/ml
11) N-OCTACOSANE	17.029	9359886	50.368 ug/ml
12) N-DOTRIACONTANE	18.753	8747148	48.031 ug/ml
14) N-HEXATRIACONTANE	20.278	8269758	46.165 ug/ml
15) N-TETRACONTANE	22.013	7688823	49.331 ug/ml
16) N-TETRATETRAOCTANE	24.961	6988034	59.594 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE121318\
Data File : FE029330.D
Signal(s) : FID1B.CH
Acq On : 13 Dec 2018 19:09
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 14 00:40:32 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE120718.M
Quant Title :
QLast Update : Fri Dec 07 23:47:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

