

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE121423\
 Data File : FE045836.D
 Signal(s) : FID1B.ch
 Acq On : 14 Dec 2023 08:04
 Operator : YP\AJ
 Sample : 1760 TRPH STD
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 1760 TRPH STD

Integration File: autoint1.e
 Quant Time: Dec 14 11:17:59 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE121423.M
 Quant Title :
 QLast Update : Thu Dec 14 11:14:22 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.934	2505085	19.512 ug/ml
Target Compounds			
1) N-OCTANE	1.967	21065231	157.795 ug/ml
2) N-NONANE	3.229	21446854	156.924 ug/ml
3) N-DECANE	4.490	32913956	235.134 ug/ml
4) N-UNDECANE	5.630	33172312	235.481 ug/ml
5) N-DODECANE	6.661	33075007	234.570 ug/ml
6) N-TETRADECANE	8.488	33105942	234.151 ug/ml
7) N-HEXADECANE	10.094	27291707	196.262 ug/ml
8) N-OCTADECANE	11.531	13435611	97.447 ug/ml
9) N-EICOSANE	12.836	5102030	38.674 ug/ml
10) N-TETRACOSANE	15.140	4883944	38.544 ug/ml
11) N-OCTACOSANE	17.110	2348201	19.342 ug/ml
12) N-DOTRIACONTANE	18.832	2353143	19.669 ug/ml
14) N-HEXATRIACONTANE	20.359	2314820	19.756 ug/ml
15) N-TETRACONTANE	21.980	1752950	20.637 ug/ml
16) N-TETRTETRACOOCTANE	24.706	971370	23.798 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE121423\
Data File : FE045836.D
Signal(s) : FID1B.ch
Acq On : 14 Dec 2023 08:04
Operator : YP\AJ
Sample : 1760 TRPH STD
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
1760 TRPH STD

Integration File: autoint1.e
Quant Time: Dec 14 11:17:59 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE121423.M
Quant Title :
QLast Update : Thu Dec 14 11:14:22 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

