

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE121525AL\  
Data File : FE057336.D  
Signal(s) : FID1B.ch  
Acq On : 15 Dec 2025 15:23  
Operator : YP\AJ  
Sample : Q3854-02  
Misc :  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
ETGI-360-E2

Integration File: sample.E  
Quant Time: Dec 16 03:36:57 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 112725.M  
Quant Title : GC Extractables  
QLast Update : Mon Dec 01 08:04:59 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.927 | 7472990    | 34.055 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 68.11%       |
| 12) S 1-chlorooctadecane (S... | 13.364 | 4911218    | 29.510 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 59.02%       |

Target Compounds

-----

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_E\Data\FE121525AL\  
Data File : FE057336.D  
Signal(s) : FID1B.ch  
Acq On : 15 Dec 2025 15:23  
Operator : YP\AJ  
Sample : Q3854-02  
Misc :  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
FID\_E  
ClientSampleId :  
ETGI-360-E2

Integration File: sample.E  
Quant Time: Dec 16 03:36:57 2025  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_E\methods\Aliphatic EPH 112725.M  
Quant Title : GC Extractables  
QLast Update : Mon Dec 01 08:04:59 2025  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.927 min  
Delta R.T.: -0.013 min  
Response: 7472990  
Conc: 34.05 ug/ml

Instrument :  
FID\_E  
ClientSampleId :  
ETGI-360-E2



#12 1-chlorooctadecane (SURR)

R.T.: 13.364 min  
Delta R.T.: -0.015 min  
Response: 4911218  
Conc: 29.51 ug/ml