

rteres

Area Percent Report

Instrument :
FID_E
LabSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
12/19/2019 10:34:12 AM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE121819\
Data File : FE030997.D
Signal(s) : FID1B.CH
Acq On : 18 Dec 2019 11:31
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE112619.M
Title :

Signal : FID1B.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.341	2.295	2.495	BV	2084871	18244899	100.00%	6.700%
2	4.381	4.328	4.492	BB	1781821	18025832	98.80%	6.620%
3	6.394	6.342	6.512	BB	1651808	18021522	98.78%	6.618%
4	8.191	8.145	8.308	BB	1526243	17660299	96.80%	6.485%
5	9.794	9.742	9.905	BB	1426052	17509875	95.97%	6.430%
6	11.239	11.192	11.345	BB	1315275	17114190	93.80%	6.285%
7	12.551	12.498	12.658	BB	1265237	17061297	93.51%	6.265%
8	13.753	13.702	13.862	BB	1191879	16735068	91.72%	6.146%
9	14.654	14.602	14.755	BB	995231	14318658	78.48%	5.258%
10	14.861	14.765	14.968	BB	1214698	16471738	90.28%	6.049%
11	15.882	15.715	15.982	BB	1120600	16170186	88.63%	5.938%
12	16.834	16.745	16.935	BB	1117331	15858446	86.92%	5.824%
13	17.723	17.668	17.822	BB	1037891	15383169	84.31%	5.649%
14	18.558	18.495	18.608	M	912021	14005105	76.76%	5.143%
15	19.342	19.282	19.387	M	810301	12456071	68.27%	4.574%
16	20.082	20.019	20.129	M	673750	10531795	57.72%	3.868%
17	20.814	20.746	20.881	M	497542	8882171	48.68%	3.262%
18	21.715	21.640	21.808	M	315353	7858256	43.07%	2.886%
Sum of corrected areas:						272308575		

FE112619.M Thu Dec 19 08:01:10 2019