

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE122118\
 Data File : FE029360.D
 Signal(s) : FID1B.CH
 Acq On : 22 Dec 2018 00:35
 Operator : AJ\SJ
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Dec 22 01:26:51 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE120718.M
 Quant Title :
 QLast Update : Fri Dec 07 23:47:48 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
13) S TETRACOSANE-d50 (SURR...	14.852	10345965	53.991 ug/ml
Target Compounds			
1) N-OCTANE	2.528	69143670	396.931 ug/ml
2) N-NONANE	3.532	75897073	435.540 ug/ml
3) N-DECANE	4.607	117166940	450.665 ug/ml
4) N-UNDECANE	5.644	123209114	678.599 ug/ml
5) N-DODECANE	6.629	121706320	675.563 ug/ml
6) N-TETRADECANE	8.433	120042899	671.556 ug/ml
7) N-HEXADECANE	10.036	102330590	553.828 ug/ml
8) N-OCTADECANE	11.463	51484160	277.927 ug/ml
9) N-EICOSANE	12.757	20683675	114.211 ug/ml
10) N-TETRACOSANE	15.069	20219078	108.732 ug/ml
11) N-OCTACOSANE	17.035	9845652	52.982 ug/ml
12) N-DOTRIACONTANE	18.761	9334447	51.256 ug/ml
14) N-HEXATRIACONTANE	20.289	8850035	49.405 ug/ml
15) N-TETRACONTANE	22.030	8321271	53.389 ug/ml
16) N-TETRATETRAOCTANE	24.992	7546005	64.353 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE122118\
Data File : FE029360.D
Signal(s) : FID1B.CH
Acq On : 22 Dec 2018 00:35
Operator : AJ\SJ
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Dec 22 01:26:51 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE120718.M
Quant Title :
QLast Update : Fri Dec 07 23:47:48 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

