

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE122319\
 Data File : FE031032.D
 Signal(s) : FID1B.CH
 Acq On : 23 Dec 2019 12:19
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Ankita
 12/24/2019 8:48:52 AM

Integration File: autoint1.e
 Quant Time: Dec 24 01:05:45 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE112619.M
 Quant Title :
 QLast Update : Tue Nov 26 14:51:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.639	15926234	53.091 ug/ml
Target Compounds			
1) N-OCTANE	2.340	19064909	51.934 ug/ml
2) N-DECANE	4.377	20303148	55.189 ug/ml
3) N-DODECANE	6.387	20366451	55.476 ug/ml
4) N-TETRADECANE	8.182	19851539	54.269 ug/ml
5) N-HEXADECANE	9.785	19593938	53.587 ug/ml
6) N-OCTADECANE	11.229	19152673	53.449 ug/ml
7) N-EICOSANE	12.540	19020668	53.229 ug/ml
8) N-DOCOSANE	13.741	18595729	52.963 ug/ml
10) N-TETRACOSANE	14.846	18225377	52.110 ug/ml
11) N-HEXACOSANE	15.867	17867413	52.164 ug/ml
12) N-OCTACOSANE	16.818	17581039	51.763 ug/ml
13) N-TRIACONTANE	17.707	17176088	51.214 ug/ml
14) N-DOTRIACONTANE	18.540	15694438	48.525 ug/ml
15) N-TETRATRIACONTANE	19.324	13268726	42.246 ug/mlm
16) N-HEXATRIACONTANE	20.061	9627837	32.372 ug/mlm
17) N-OCTATRIACONTANE	20.788	6355631	22.947 ug/mlm
18) N-TETRACONTANE	21.680	4141374	14.638 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE122319\
Data File : FE031032.D
Signal(s) : FID1B.CH
Acq On : 23 Dec 2019 12:19
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
12/24/2019 8:48:52 AM

Integration File: autoint1.e
Quant Time: Dec 24 01:05:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\FE112619.M
Quant Title :
QLast Update : Tue Nov 26 14:51:59 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

