

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/27/2023
Supervised By :Ankita Jodhani 12/28/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE
Data File : FE045973.D
Signal(s) : FID1B.ch
Acq On : 26 Dec 2023 19:52
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 113023.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.226	3.201	3.290	BB	333789	3357102	85.07%	4.647%
2	4.477	4.436	4.546	BB	336064	3420412	86.67%	4.735%
3	6.179	6.139	6.282	BB	355363	3665453	92.88%	5.074%
4	6.645	6.599	6.709	BB	345443	3457706	87.62%	4.786%
5	7.275	7.231	7.391	BB	314436	3543826	89.80%	4.905%
6	8.472	8.426	8.554	BB	311087	3412554	86.47%	4.724%
7	10.082	10.034	10.155	BB	300904	3448781	87.39%	4.774%
8	11.528	11.432	11.588	BB	289491	3416873	86.58%	4.730%
9	11.841	11.796	11.912	BB	335977	3946319	100.00%	5.462%
10	12.839	12.792	12.895	BB	268201	3255464	82.49%	4.506%
11	13.285	13.237	13.342	BB	216270	2890932	73.26%	4.002%
12	13.452	13.403	13.504	BB	249071	3171542	80.37%	4.390%
13	14.040	13.964	14.103	BB	234478	3123783	79.16%	4.324%
14	15.146	15.096	15.209	BB	214324	3038578	77.00%	4.206%
15	16.170	16.114	16.232	BB	188467	2977685	75.45%	4.122%
16	17.123	17.066	17.189	BB	185824	2974301	75.37%	4.117%
17	18.014	17.821	18.084	BB	175664	3159592	80.06%	4.374%
18	18.852	18.493	18.922	BV	161288	3248027	82.31%	4.496%
19	19.637	19.570	19.707	BV	150142	3110410	78.82%	4.305%
20	20.383	20.337	20.448	PV	137115	3109002	78.78%	4.303%
21	21.115	21.050	21.189	BBA	113694	3175939	80.48%	4.396%
22	22.025	21.953	22.156	VV	81834	3339676	84.63%	4.623%
Sum of corrected areas:						72243957		

Aliphatic EPH 113023.M Wed Dec 27 03:30:24 2023