

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE122822AL\
 Data File : FE039508.D
 Signal(s) : FID1B.ch
 Acq On : 28 Dec 2022 08:13
 Operator : YP\AJ
 Sample : MECL2
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 FID_E
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Dec 28 11:13:16 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 121722.M
 Quant Title : GC Extractables
 QLast Update : Sat Dec 17 15:09:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds

10) T	n-Eicosane (C20)	12.904	99804	0.731 ug/ml
11) T	n-Heneicosane (C21)	13.521	74968	0.550 ug/ml
13) T	n-Docosane (C22)	14.111	93369	0.690 ug/ml
14) T	n-Tetracosane (C24)	15.196	160499	1.196 ug/ml
15) T	n-Hexacosane (C26)	16.237	58811	0.452 ug/ml
16) T	n-Octacosane (C28)	17.187	68150	0.530 ug/ml
17) T	n-Tricontane (C30)	18.076	48655	0.378 ug/ml
18) T	n-Dotriacontane (C32)	18.908	109891	0.849 ug/ml
19) T	n-Tetratriacontane (C34)	19.692	70862	0.562 ug/ml
20) T	n-Hexatriacontane (C36)	20.431	73417	0.607 ug/ml
21) T	n-Octatriacontane (C38)	21.169	94655	0.766 ug/ml
22) T	n-Tetracontane (C40)	22.076	279365	2.225 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE122822AL\
Data File : FE039508.D
Signal(s) : FID1B.ch
Acq On : 28 Dec 2022 08:13
Operator : YP\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_E
ClientSampleId :

Integration File: autoint1.e
Quant Time: Dec 28 11:13:16 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 121722.M
Quant Title : GC Extractables
QLast Update : Sat Dec 17 15:09:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 0.000 min
Exp R.T.: 11.950 min
Response: 0
Conc: N.D.

Instrument :
FID_E
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.904 min
Delta R.T.: -0.033 min
Response: 99804
Conc: 0.73 ug/ml




#11 n-Heneicosane (C21)

R.T.: 13.521 min
Delta R.T.: -0.029 min
Response: 74968
Conc: 0.55 ug/ml



#12 1-chlorooctadecane (SURR)


R.T.: 0.000 min
Exp R.T.: 13.383 min
Response: 0
Conc: N.D.



#13 n-Docosane (C22)


R.T.: 14.111 min
Delta R.T.: -0.027 min
Response: 93369
Conc: 0.69 ug/ml

Instrument :
FID_E
ClientSampleId :




#14 n-Tetracosane (C24)

R.T.: 15.196 min
Delta R.T.: -0.043 min
Response: 160499
Conc: 1.20 ug/ml




#15 n-Hexacosane (C26)

R.T.: 16.237 min
Delta R.T.: -0.025 min
Response: 58811
Conc: 0.45 ug/ml



#16 n-Octacosane (C28)


R.T.: 17.187 min
Delta R.T.: -0.025 min
Response: 68150
Conc: 0.53 ug/ml



#17 n-Tricontane (C30)


R.T.: 18.076 min
Delta R.T.: -0.025 min
Response: 48655
Conc: 0.38 ug/ml

Instrument :
FID_E
ClientSampleId :




#18 n-Dotriacontane (C32)

R.T.: 18.908 min
Delta R.T.: -0.026 min
Response: 109891
Conc: 0.85 ug/ml




#19 n-Tetratriacontane (C34)

R.T.: 19.692 min
Delta R.T.: -0.026 min
Response: 70862
Conc: 0.56 ug/ml



#20 n-Hexatriacontane (C36)


R.T.: 20.431 min
Delta R.T.: -0.027 min
Response: 73417
Conc: 0.61 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 21.169 min
Delta R.T.: -0.031 min
Response: 94655
Conc: 0.77 ug/ml

Instrument :
FID_E
ClientSampleId :



#22 n-Tetracontane (C40)

R.T.: 22.076 min
Delta R.T.: -0.047 min
Response: 279365
Conc: 2.22 ug/ml