

Instrument :
FID_E
LabSampleId :
20 PPM ALIPHATIC HC STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/29/2023
Supervised By :Ankita Jodhani 12/29/2023

rteres

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE
Data File : FE045997.D
Signal(s) : FID1B.ch
Acq On : 28 Dec 2023 05:59
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\Aliphatic EPH 113023.M
Title : GC Extractables

Signal : FID1B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.223	3.202	3.293	BB	336890	3421971	83.15%	4.567%
2	4.474	4.434	4.544	BB	346565	3517147	85.46%	4.694%
3	6.175	6.138	6.287	BB	350293	3768671	91.57%	5.029%
4	6.642	6.600	6.707	BB	348227	3547760	86.20%	4.734%
5	7.272	7.227	7.397	BV	317260	3636576	88.36%	4.853%
6	8.468	8.422	8.561	BB	322788	3515628	85.42%	4.691%
7	10.079	10.033	10.165	BB	303726	3562529	86.56%	4.754%
8	11.525	11.480	11.584	BB	306175	3539046	85.99%	4.723%
9	11.838	11.796	11.908	BB	351820	4115503	100.00%	5.492%
10	12.837	12.787	12.900	BB	268511	3396055	82.52%	4.532%
11	13.283	13.236	13.348	BB	247447	3017869	73.33%	4.027%
12	13.450	13.402	13.507	BB	251871	3315292	80.56%	4.424%
13	14.037	13.959	14.105	BB	247538	3264039	79.31%	4.356%
14	15.144	15.092	15.210	BB	223575	3178636	77.24%	4.242%
15	16.167	16.114	16.241	BB	212870	3121011	75.84%	4.165%
16	17.120	17.059	17.198	BB	193886	3124179	75.91%	4.169%
17	18.011	17.956	18.093	BB	185533	3274261	79.56%	4.369%
18	18.847	18.775	18.939	BB	180144	3329084	80.89%	4.443%
19	19.634	19.584	19.725	BB	167352	3270530	79.47%	4.364%
20	20.377	20.320	20.460	BBA	157960	3266512	79.37%	4.359%
21	21.111	21.050	21.190	BBA	120348	3295500	80.08%	4.398%
22	22.017	21.942	22.180	VV	88107	3458522	84.04%	4.615%
Sum of corrected areas:						74936321		

Aliphatic EPH 113023.M Fri Dec 29 01:31:13 2023