

Method Path : P:\HPCHEM1\FID_E\methods\
 Method File : C44-FE012918.M
 Title :
 Last Update : Mon Jan 29 14:42:51 2018
 Response Via : Initial Calibration

Calibration Files

= S1 =FE025951.D S2 =FE025952.D
 S3 =FE025953.D S4 =FE025954.D S5 =FE025955.D

	Compound		S1	S2	S3	S4	S5	Avg		%RSD
1)	N-OCTANE	0.946	1.041	0.990	0.968	1.044		0.998	E5	4.36
2)	N-NONANE	1.001	1.092	1.035	1.010	1.083		1.044	E5	3.98
3)	N-DECANE	1.572	1.700	1.612	1.573	1.678		1.627	E5	3.66
4)	N-UNDECANE	1.147	1.233	1.172	1.142	1.207		1.180	E5	3.32
5)	N-DODECANE	1.182	1.266	1.210	1.182	1.236		1.215	E5	2.97
6)	N-TETRADECANE	1.250	1.332	1.299	1.283	1.319		1.297	E5	2.46
7)	N-HEXADECANE	1.348	1.421	1.419	1.426	1.457		1.414	E5	2.85
8)	N-OCTADECANE	1.410	1.463	1.495	1.536	1.579		1.496	E5	4.35
9)	N-EICOSANE	1.465	1.475	1.529	1.598	1.615		1.536	E5	4.47
10)	N-TETRACOSANE	1.410	1.265	1.328	1.466	1.365		1.367	E5	5.61
11)	N-OCTACOSANE	1.145	1.073	1.095	1.190	1.133		1.127	E5	4.04
12)	N-DOTRIACONTANE	1.100	1.141	1.157	1.230	1.277		1.181	E5	6.05
13) S	TETRACOSANE-d...	1.653	1.499	1.616	1.719	1.673		1.632	E5	5.09
14)	N-HEXATRIACON...	1.089	0.969	1.160	1.217	1.193		1.126	E5	8.86
15)	N-TETRACONTANE	0.964	1.294	1.193	1.313	1.171		1.187	E5	11.72
16)	N-TETRTETRACO...	0.807	1.147	1.152	1.411	1.073		1.118	E5	19.34

(#) = Out of Range