

Method Path : P:\HPCHEM1\FID_E\methods\
 Method File : C44-FE022618.M
 Title :
 Last Update : Tue Feb 27 00:52:02 2018
 Response Via : Initial Calibration

Calibration Files

= S1 =FE026307.D S2 =FE026308.D
 S3 =FE026309.D S4 =FE026310.D S5 =FE026311.D

	Compound	S1	S2	S3	S4	S5	Avg		%RSD
1)	N-OCTANE	1.021	0.966	0.962	0.993	0.966	0.982	E5	2.58
2)	N-NONANE	1.055	1.004	0.991	1.013	0.985	1.010	E5	2.74
3)	N-DECANE	1.620	1.548	1.528	1.554	1.518	1.554	E5	2.56
4)	N-UNDECANE	1.156	1.112	1.098	1.110	1.085	1.112	E5	2.41
5)	N-DODECANE	1.166	1.129	1.121	1.127	1.098	1.128	E5	2.16
6)	N-TETRADECANE	1.187	1.169	1.178	1.181	1.152	1.173	E5	1.18
7)	N-HEXADECANE	1.238	1.238	1.267	1.266	1.236	1.249	E5	1.29
8)	N-OCTADECANE	1.266	1.283	1.330	1.341	1.329	1.310	E5	2.53
9)	N-EICOSANE	1.292	1.322	1.370	1.376	1.345	1.341	E5	2.58
10)	N-TETRACOSANE	1.267	1.304	1.323	1.351	1.292	1.307	E5	2.41
11)	N-OCTACOSANE	1.216	1.223	1.225	1.292	1.216	1.234	E5	2.63
12)	N-DOTRIACONTANE	1.141	1.152	1.170	1.218	1.218	1.180	E5	3.07
13) S	TETRACOSANE-d...	1.521	1.559	1.579	1.604	1.549	1.562	E5	2.00
14)	N-HEXATRIACON...	1.170	1.167	1.190	1.188	1.182	1.180	E5	0.88
15)	N-TETRACONTANE	0.951	0.977	1.068	0.905	0.999	0.980	E5	6.16
16)	N-TETRTETRACO...	6.748	7.304	8.902	5.738	7.512	7.241	E4	15.96

(#) = Out of Range