

Method Path : P:\HPCHEM1\FID_E\methods\
 Method File : C44-FE032318.M
 Title :
 Last Update : Fri Mar 23 12:17:01 2018
 Response Via : Initial Calibration

Calibration Files

= S1 =FE026460.D S2 =FE026461.D
 S3 =FE026462.D S4 =FE026463.D S5 =FE026464.D

	Compound		S1	S2	S3	S4	S5	Avg		%RSD
1)	N-OCTANE	8.237	9.051	9.486	8.381	8.946		8.820	E4	5.80
2)	N-NONANE	8.752	9.499	9.872	8.754	9.285		9.233	E4	5.26
3)	N-DECANE	1.393	1.493	1.536	1.364	1.433		1.444	E5	4.89
4)	N-UNDECANE	1.036	1.101	1.123	0.999	1.038		1.059	E5	4.80
5)	N-DODECANE	1.090	1.153	1.171	1.047	1.075		1.107	E5	4.77
6)	N-TETRADECANE	1.200	1.268	1.293	1.172	1.190		1.225	E5	4.30
7)	N-HEXADECANE	1.340	1.416	1.458	1.346	1.378		1.388	E5	3.57
8)	N-OCTADECANE	1.450	1.526	1.586	1.487	1.546		1.519	E5	3.45
9)	N-EICOSANE	1.555	1.624	1.692	1.609	1.686		1.633	E5	3.50
10)	N-TETRACOSANE	1.612	1.627	1.688	1.693	1.739		1.672	E5	3.12
11)	N-OCTACOSANE	1.486	1.389	1.387	1.490	1.514		1.453	E5	4.16
12)	N-DOTRIACONTANE	1.296	1.283	1.338	1.413	1.468		1.360	E5	5.81
13) S	TETRACOSANE-d...	1.765	1.792	1.856	1.861	1.899		1.835	E5	2.99
14)	N-HEXATRIACON...	1.291	1.252	1.314	1.385	1.368		1.322	E5	4.16
15)	N-TETRACONTANE	1.063	0.936	1.101	1.245	0.980		1.065	E5	11.27
16)	N-TETRTETRACO...	7.065	5.265	7.031	8.282	5.273		6.583	E4	19.76

(#) = Out of Range