

Data Path : P:\HPCHEM1\FID E\Data\FE040618\
 Data File : FE026551.D
 Signal(s) : FID1B.CH
 Acq On : 06 Apr 2018 13:00
 Operator : UA/AJ
 Sample : J2212-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Apr 06 23:41:21 2018
 Quant Method : P:\HPCHEM1\FID_E\methods\C44-FE032318.M
 Quant Title :
 QLast Update : Fri Mar 23 12:17:01 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
13) S	TETRACOSANE-d50 (SURR...	15.329	3788952	20.651 ug/ml
Target Compounds				
4)	N-UNDECANE	6.094	139108	1.313 ug/ml
6)	N-TETRADECANE	8.845	852987	6.965 ug/ml
7)	N-HEXADECANE	10.439	719244	5.183 ug/ml
8)	N-OCTADECANE	11.891	554563	3.651 ug/ml
9)	N-EICOSANE	13.224	3760040	23.025 ug/ml
10)	N-TETRACOSANE	15.555	4038885	24.160 ug/ml
11)	N-OCTACOSANE	17.533	4026736	27.710 ug/ml
12)	N-DOTRIACONTANE	19.245	2358569	17.348 ug/ml
14)	N-HEXATRIACONTANE	20.810	3751537	28.375 ug/ml
15)	N-TETRACONTANE	22.908	242880	2.280 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : P:\HPCHEM1\FID E\Data\FE040618\
Data File : FE026551.D
Signal(s) : FID1B.CH
Acq On : 06 Apr 2018 13:00
Operator : UA/AJ
Sample : J2212-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Apr 06 23:41:21 2018
Quant Method : P:\HPCHEM1\FID_E\methods\C44-FE032318.M
Quant Title :
QLast Update : Fri Mar 23 12:17:01 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um






#4 N-UNDECANE

R.T.: 6.094 min
Delta R.T.: 0.061 min
Response: 139108
Conc: 1.31 ug/ml




#6 N-TETRADECANE

R.T.: 8.845 min
Delta R.T.: 0.006 min
Response: 852987
Conc: 6.96 ug/ml




#7 N-HEXADECANE

R.T.: 10.439 min
Delta R.T.: -0.014 min
Response: 719244
Conc: 5.18 ug/ml




#8 N-OCTADECANE

R.T.: 11.891 min
Delta R.T.: -0.011 min
Response: 554563
Conc: 3.65 ug/ml




#9 N-EICOSANE

R.T.: 13.224 min
Delta R.T.: 0.006 min
Response: 3760040
Conc: 23.02 ug/ml




#10 N-TETRACOSANE

R.T.: 15.555 min
Delta R.T.: 0.018 min
Response: 4038885
Conc: 24.16 ug/ml




#11 N-OCTACOSANE

R.T.: 17.533 min
Delta R.T.: 0.016 min
Response: 4026736
Conc: 27.71 ug/ml




#12 N-DOTRIACONTANE

R.T.: 19.245 min
Delta R.T.: -0.001 min
Response: 2358569
Conc: 17.35 ug/ml




#13 TETRACOSANE-d50 (SURROGATE)

R.T.: 15.329 min
Delta R.T.: 0.000 min
Response: 3788952
Conc: 20.65 ug/ml



#14 N-HEXATRIACONTANE

R.T.: 20.810 min
Delta R.T.: 0.006 min
Response: 3751537
Conc: 28.38 ug/ml



#15 N-TETRACONTANE

R.T.: 22.908 min
Delta R.T.: -0.001 min
Response: 242880
Conc: 2.28 ug/ml