

Method Path : P:\HPCHEM1\FID_E\methods\
 Method File : FE040918.M
 Title :
 Last Update : Mon Apr 09 14:09:36 2018
 Response Via : Initial Calibration

Calibration Files

50 =FE026564.D 100 =FE026563.D 20 =FE026565.D
 10 =FE026566.D 5 =FE026567.D

	Compound	50	100	20	10	5	Avg		%RSD
1)	N-OCTANE	1.599	1.572	1.617	1.603	1.550	1.588	E5	1.70
2)	N-DECANE	1.734	1.723	1.746	1.743	1.639	1.717	E5	2.59
3)	N-DODECANE	1.909	1.933	1.901	1.896	1.752	1.878	E5	3.82
4)	N-TETRADECANE	2.100	2.137	2.068	2.060	1.886	2.050	E5	4.72
5)	N-HEXADECANE	2.245	2.257	2.226	2.246	2.102	2.215	E5	2.91
6)	N-OCTADECANE	2.320	2.312	2.324	2.372	2.367	2.339	E5	1.21
7)	N-EICOSANE	2.301	2.294	2.314	2.351	2.397	2.331	E5	1.84
8)	N-DOCOSANE	2.233	2.250	2.293	2.324	2.392	2.298	E5	2.75
9) S	TETRACOSANE-d50 (S...	1.867	1.913	1.960	1.994	2.034	1.954	E5	3.38
10)	N-TETRACOSANE	2.092	2.159	2.239	2.296	2.327	2.223	E5	4.36
11)	N-HEXACOSANE	1.862	1.988	2.108	2.242	2.194	2.079	E5	7.45
12)	N-OCTACOSANE	1.631	1.729	1.845	2.074	1.932	1.842	E5	9.37
13)	N-TRIACONTANE	1.608	1.629	1.683	1.876	1.773	1.714	E5	6.48
14)	N-DOTRIACONTANE	1.683	1.701	1.752	1.941	1.876	1.791	E5	6.32
15)	N-TETRATRIACONTANE	1.763	1.766	1.826	1.949	1.923	1.846	E5	4.71
16)	N-HEXATRIACONTANE	1.769	1.763	1.814	1.923	1.900	1.834	E5	4.04
17)	N-OCTATRIACONTANE	1.662	1.660	1.700	1.778	1.761	1.712	E5	3.22
18)	N-TETRACONTANE	1.692	1.686	1.724	1.783	1.760	1.729	E5	2.43

(#) = Out of Range