

Method Path : P:\HPCHEM1\FID E\methods\
 Method File : C44-FE042118.M
 Title :
 Last Update : Sat Apr 21 01:24:29 2018
 Response Via : Initial Calibration

Calibration Files

= S1 =FE026698.D S2 =FE026699.D
 S3 =FE026700.D S4 =FE026701.D S5 =FE026702.D

	Compound	S1	S2	S3	S4	S5	Avg	%RSD
1)	N-OCTANE	1.061	1.086	1.082	1.139	1.296	1.133 E5	8.46
2)	N-NONANE	1.131	1.131	1.120	1.182	1.330	1.179 E5	7.44
3)	N-DECANE	1.788	1.757	1.730	1.826	2.028	1.826 E5	6.50
4)	N-UNDECANE	1.310	1.273	1.252	1.322	1.450	1.322 E5	5.86
5)	N-DODECANE	1.349	1.307	1.287	1.365	1.479	1.357 E5	5.51
6)	N-TETRADECANE	1.410	1.380	1.373	1.477	1.584	1.445 E5	6.09
7)	N-HEXADECANE	1.488	1.488	1.496	1.639	1.780	1.578 E5	8.23
8)	N-OCTADECANE	1.512	1.561	1.576	1.740	1.951	1.668 E5	10.79
9)	N-EICOSANE	1.509	1.619	1.624	1.761	2.042	1.711 E5	12.01
10)	N-TETRACOSANE	1.403	1.590	1.589	1.605	1.948	1.627 E5	12.16
11)	N-OCTACOSANE	1.408	1.506	1.556	1.593	1.866	1.586 E5	10.81
12)	N-DOTRIACONTANE	1.378	1.429	1.545	1.589	1.866	1.561 E5	12.19
13) S	TETRACOSANE-d...	1.556	1.748	1.745	1.769	2.143	1.792 E5	11.96
14)	N-HEXATRIACON...	1.366	1.440	1.509	1.517	1.727	1.512 E5	8.93
15)	N-TETRACONTANE	1.100	1.282	1.221	1.263	1.369	1.247 E5	7.86
16)	N-TETRTETRACO...	0.837	1.129	0.985	1.043	1.044	1.008 E5	10.74

(#) = Out of Range