

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE070518\
 Data File : FE028063.D
 Signal(s) : FID1B.CH
 Acq On : 05 Jul 2018 19:35
 Operator : AJ\MA
 Sample : 4400 PPM TRPH STD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 06 09:47:05 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE062618.M
 Quant Title :
 QLast Update : Tue Jun 26 18:10:02 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20m x 0.18mm x 0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
13) S	TETRACOSANE-d50 (SURR...	15.037	9651702	28.924 ug/ml
Target Compounds				
1)	N-OCTANE	2.623	48412540	400.960 ug/ml
2)	N-NONANE	3.661	52383222	394.764 ug/ml
3)	N-DECANE	4.750	83635747	387.811 ug/ml
4)	N-UNDECANE	5.796	89602898	567.508 ug/ml
5)	N-DODECANE	6.787	93599848	548.811 ug/ml
6)	N-TETRADECANE	8.599	100348313	504.044 ug/ml
7)	N-HEXADECANE	10.208	87995279	381.834 ug/ml
8)	N-OCTADECANE	11.640	45042712	174.660 ug/ml
9)	N-EICOSANE	12.939	17998008	65.060 ug/ml
10)	N-TETRACOSANE	15.251	16405009	57.797 ug/ml
11)	N-OCTACOSANE	17.218	6910279	25.921 ug/ml
12)	N-DOTRIACONTANE	18.943	5921286	25.549 ug/ml
14)	N-HEXATRIACONTANE	20.471	5024217	27.212 ug/ml
15)	N-TETRACONTANE	22.342	5124783	30.357 ug/ml
16)	N-TETRTETRACOOCTANE	25.573	4939578	33.516 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_E\Data\FE070518\
Data File : FE028063.D
Signal(s) : FID1B.CH
Acq On : 05 Jul 2018 19:35
Operator : AJ\MA
Sample : 4400 PPM TRPH STD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 06 09:47:05 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_E\methods\C44-FE062618.M
Quant Title :
QLast Update : Tue Jun 26 18:10:02 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20m x 0.18mm x 0.18um

