

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG010419\
 Data File : FG003373.D
 Signal(s) : FID1A.ch
 Acq On : 04 Jan 2019 12:48
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 04 21:44:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121318.M
 Quant Title :
 QLast Update : Thu Dec 13 00:59:54 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.699	6333463	51.065 ug/ml
Target Compounds			
1) N-OCTANE	2.300	7166311	54.228 ug/ml
2) N-DECANE	4.405	7712995	57.487 ug/ml
3) N-DODECANE	6.431	7652897	56.251 ug/ml
4) N-TETRADECANE	8.231	7824076	56.694 ug/ml
5) N-HEXADECANE	9.836	7807901	55.730 ug/ml
6) N-OCTADECANE	11.282	7702988	54.370 ug/ml
7) N-EICOSANE	12.595	7466548	52.852 ug/ml
8) N-DOCOSANE	13.797	7218622	51.324 ug/ml
10) N-TETRACOSANE	14.905	6990624	49.943 ug/ml
11) N-HEXACOSANE	15.929	6821021	48.939 ug/ml
12) N-OCTACOSANE	16.886	6734464	48.193 ug/ml
13) N-TRIACONTANE	17.777	6709353	47.312 ug/ml
14) N-DOTRIACONTANE	18.616	6510618	47.189 ug/ml
15) N-TETRATRIACONTANE	19.406	6363336	48.290 ug/mlm
16) N-HEXATRIACONTANE	20.156	6135493	48.389 ug/ml
17) N-OCTATRIACONTANE	20.920	5578142	49.763 ug/ml
18) N-TETRACONTANE	21.861	5472605	47.989 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG010419\
Data File : FG003373.D
Signal(s) : FID1A.ch
Acq On : 04 Jan 2019 12:48
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 04 21:44:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG121318.M
Quant Title :
QLast Update : Thu Dec 13 00:59:54 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

