

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG010819\
 Data File : FG003415.D
 Signal(s) : FID1A.ch
 Acq On : 08 Jan 2019 20:38
 Operator : AJ\SJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 09 08:14:24 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:05:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.483	12627075	100.511 ug/ml
Target Compounds			
1) N-OCTANE	2.116	13857211	103.664 ug/ml
2) N-DECANE	4.207	13986169	103.241 ug/ml
3) N-DODECANE	6.232	14312872	104.136 ug/ml
4) N-TETRADECANE	8.027	14649866	104.800 ug/ml
5) N-HEXADECANE	9.628	14797379	104.053 ug/ml
6) N-OCTADECANE	11.071	14836903	103.185 ug/ml
7) N-EICOSANE	12.382	14621343	102.018 ug/ml
8) N-DOCOSANE	13.582	14400369	101.108 ug/ml
10) N-TETRACOSANE	14.687	14088865	99.547 ug/ml
11) N-HEXACOSANE	15.709	13858974	98.510 ug/ml
12) N-OCTACOSANE	16.660	13580419	96.538 ug/ml
13) N-TRIACONTANE	17.548	13264945	93.256 ug/ml
14) N-DOTRIACONTANE	18.383	12616105	91.452 ug/ml
15) N-TETRATRIACONTANE	19.168	12118640	92.195 ug/ml
16) N-HEXATRIACONTANE	19.913	11546152	91.738 ug/ml
17) N-OCTATRIACONTANE	20.631	10275982	92.268 ug/ml
18) N-TETRACONTANE	21.487	10426664	92.343 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG010819\
Data File : FG003415.D
Signal(s) : FID1A.ch
Acq On : 08 Jan 2019 20:38
Operator : AJ\SJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 09 08:14:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:05:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

