

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG010819\
 Data File : FG003419.D
 Signal(s) : FID1A.ch
 Acq On : 08 Jan 2019 22:39
 Operator : AJ\SJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 09 08:20:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:05:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.472	673332	5.360 ug/ml
Target Compounds			
1) N-OCTANE	2.120	656248	4.909 ug/ml
2) N-DECANE	4.203	611457	4.514 ug/ml
3) N-DODECANE	6.225	563165	4.097 ug/ml
4) N-TETRADECANE	8.020	601971	4.306 ug/ml
5) N-HEXADECANE	9.620	674105	4.740 ug/ml
6) N-OCTADECANE	11.062	719609	5.005 ug/ml
7) N-EICOSANE	12.373	744417	5.194 ug/ml
8) N-DOCOSANE	13.571	746821	5.244 ug/ml
10) N-TETRACOSANE	14.675	746144	5.272 ug/ml
11) N-HEXACOSANE	15.696	743778	5.287 ug/ml
12) N-OCTACOSANE	16.647	748951	5.324 ug/ml
13) N-TRIACONTANE	17.534	756071	5.315 ug/ml
14) N-DOTRIACONTANE	18.369	730456	5.295 ug/ml
15) N-TETRATRIACONTANE	19.153	683172	5.197 ug/ml
16) N-HEXATRIACONTANE	19.897	614078	4.879 ug/ml
17) N-OCTATRIACONTANE	20.615	559469	5.023 ug/ml
18) N-TETRACONTANE	21.466	551027	4.880 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG010819\
Data File : FG003419.D
Signal(s) : FID1A.ch
Acq On : 08 Jan 2019 22:39
Operator : AJ\SJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 09 08:20:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:05:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

