

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011019\
 Data File : FG003446.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jan 2019 13:43
 Operator : AJ\SJ
 Sample : PB116255BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 11 02:11:24 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.473	2368984	18.258 ug/ml
Target Compounds			
1) N-OCTANE	2.117	2739564	20.316 ug/ml
2) N-DECANE	4.202	2801674	21.209 ug/ml
3) N-DODECANE	6.225	2748019	21.099 ug/ml
4) N-TETRADECANE	8.019	2732830	20.022 ug/ml
5) N-HEXADECANE	9.620	2767846	19.350 ug/ml
6) N-OCTADECANE	11.061	2795549	19.083 ug/ml
7) N-EICOSANE	12.372	2767372	18.805 ug/ml
8) N-DOCOSANE	13.571	2711991	18.576 ug/ml
10) N-TETRACOSANE	14.676	2634247	18.235 ug/ml
11) N-HEXACOSANE	15.697	2563401	17.927 ug/ml
12) N-OCTACOSANE	16.649	2510888	17.691 ug/ml
13) N-TRIACONTANE	17.537	2478293	17.590 ug/ml
14) N-DOTRIACONTANE	18.372	2457526	18.168 ug/ml
15) N-TETRATRIACONTANE	19.160	2353316	18.278 ug/ml
16) N-HEXATRIACONTANE	19.902	2274765	19.061 ug/mlm
17) N-OCTATRIACONTANE	20.622	2091453	19.494 ug/ml
18) N-TETRACONTANE	21.475	2512036	23.756 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011019\
Data File : FG003446.D
Signal(s) : FID1A.ch
Acq On : 10 Jan 2019 13:43
Operator : AJ\SJ
Sample : PB116255BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 11 02:11:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

