

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011019\
 Data File : FG003452.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jan 2019 18:55
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 11 02:12:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.473	6024994	46.435 ug/ml
Target Compounds			
1) N-OCTANE	2.116	7000543	51.915 ug/ml
2) N-DECANE	4.203	7184441	54.387 ug/ml
3) N-DODECANE	6.227	7129086	54.736 ug/ml
4) N-TETRADECANE	8.021	7184504	52.638 ug/ml
5) N-HEXADECANE	9.622	7224298	50.506 ug/ml
6) N-OCTADECANE	11.064	7218338	49.275 ug/ml
7) N-EICOSANE	12.373	7085741	48.150 ug/ml
8) N-DOCOSANE	13.572	6892177	47.209 ug/ml
10) N-TETRACOSANE	14.677	6681507	46.251 ug/ml
11) N-HEXACOSANE	15.698	6485110	45.352 ug/ml
12) N-OCTACOSANE	16.647	6343823	44.696 ug/ml
13) N-TRIACONTANE	17.539	6302682	44.734 ug/ml
14) N-DOTRIACONTANE	18.375	6162465	45.559 ug/ml
15) N-TETRATRIACONTANE	19.161	6007473	46.659 ug/ml
16) N-HEXATRIACONTANE	19.907	5875434	49.233 ug/ml
17) N-OCTATRIACONTANE	20.624	5276988	49.186 ug/ml
18) N-TETRACONTANE	21.477	5444308	51.485 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011019\
Data File : FG003452.D
Signal(s) : FID1A.ch
Acq On : 10 Jan 2019 18:55
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 11 02:12:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

