

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011419\
 Data File : FG003480.D
 Signal(s) : FID1A.ch
 Acq On : 14 Jan 2019 16:45
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 14 21:53:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
 Quant Title :
 QLast Update : Wed Jan 09 08:22:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.476	6729765	51.867 ug/ml
Target Compounds			
2) N-DECANE	4.204	8209905	62.150 ug/ml
3) N-DODECANE	6.227	8157459	62.632 ug/ml
4) N-TETRADECANE	8.022	8197094	60.057 ug/ml
5) N-HEXADECANE	9.623	8190000	57.257 ug/ml
6) N-OCTADECANE	11.065	8145197	55.602 ug/ml
7) N-EICOSANE	12.375	7950820	54.029 ug/ml
8) N-DOCOSANE	13.575	7701068	52.750 ug/ml
10) N-TETRACOSANE	14.680	7440774	51.507 ug/ml
11) N-HEXACOSANE	15.702	7226168	50.535 ug/ml
12) N-OCTACOSANE	16.654	7090655	49.958 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011419\
Data File : FG003480.D
Signal(s) : FID1A.ch
Acq On : 14 Jan 2019 16:45
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 14 21:53:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010819.M
Quant Title :
QLast Update : Wed Jan 09 08:22:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

