

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011421\
 Data File : FG007558.D
 Signal(s) : FID1A.ch
 Acq On : 14 Jan 2021 18:07
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Jan 14 20:57:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010421.M
 Quant Title :
 QLast Update : Tue Jan 05 02:18:25 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.644	6002968	50.091 ug/ml
Target Compounds			
2) N-DECANE	4.342	7593926	50.474 ug/ml
3) N-DODECANE	6.369	7626939	50.286 ug/ml
4) N-TETRADECANE	8.170	7616547	50.214 ug/ml
5) N-HEXADECANE	9.776	7507592	50.246 ug/ml
6) N-OCTADECANE	11.223	7502943	50.303 ug/ml
7) N-EICOSANE	12.537	7328122	50.280 ug/ml
8) N-DOCOSANE	13.740	7025621	50.178 ug/ml
10) N-TETRACOSANE	14.848	6868288	50.217 ug/ml
11) N-HEXACOSANE	15.873	6603189	50.277 ug/ml
12) N-OCTACOSANE	16.827	6431522	50.640 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG011421\
Data File : FG007558.D
Signal(s) : FID1A.ch
Acq On : 14 Jan 2021 18:07
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Jan 14 20:57:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG010421.M
Quant Title :
QLast Update : Tue Jan 05 02:18:25 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

